



**ARCK**

**TECHNOLOGIES**

INNOVATIONS PVT LTD

## Certificate of Compliance

- Product Description** : Culture Dish 35mm x 10mm, Pack of 5, Sterile
- Catalog Number** : 40002015
- Lot Number** : 35CD1122
- Manufacture Date** : 04/2023
- Expiry Date** : 04/2026
- Visual Attribute** : PASS
- Functional Test** : PASS
- Sterilization** : Sterilization was performed by gamma irradiation according to ISO 11137-1, the Sterilization does <SAL 10-6> establishing by ISO 1137-2
- Cytotoxicity Test** : Cytotoxicity test is conducted to qualify all material Resins using USP <87> and or ISO 10993-5 standards for Cytotoxicity and have been shown to be nontoxic.
- RNase/DNase/DNA Test** : This product has been tested and is free of any detectable RNase/DNase/DNA contamination
- MEA Test** : 1 Cell to 2 Cell within 24 Hours  
1 Cell to Expanded blastocyst within 96 hours.  
100% achieved.(Embryotech USA-AT-9805-1122)
- Pyrogen/ Endotoxin Test** : Non-pyrogenic tested according to the principles of the LAL test.  
Test Result: <0.03125EU/Device. (ELI Accession No: E2707-1122AT)

ARCK TECHNOLOGIES INNOVATIONS PVT LTD. hereby certifies that the product identified above have been inspected in compliance with product quality specification and requirements as documented.

Date: 20<sup>th</sup> May 2023

Authorized Signatory

