



ARCK
TECHNOLOGIES[™]
INNOVATIONS PVT LTD

Certificate of Compliance

Product Description : 3ml Pasteur Pipette (Individually packed-Sterile)
Catalog Number : TR003001
Lot Number : 3PCH021W / 31PCI011W
Manufacture Date : 07/2022 / 03/2023
Expiry Date : 07/2025 / 03/2026
Visual Attribute : PASS
Functional Test : PASS
Sterilization : Sterilization was performed by gamma irradiation according to ISO 11137-1, the Sterilization does <SAL 10⁻⁶> establishing by ISO 1137-2
Cytotoxicity Test : Cytotoxicity test is conducted to qualify all material Resins using USP <87> and or ISO 10993-5 standards for Cytotoxicity and have been shown to be nontoxic
RNase/DNase/DNA Test : This product has been tested and is free of any detectable RNase/DNase/DNA contamination
MEA Test : 1 Cell to 2 Cell within 24 Hours
1 Cell to Expanded blastocyst within 96 hours.
100% achieved.(Embryotech USA- AT-4077-0620)

ARCK TECHNOLOGIES INNOVATIONS PVT LTD .hereby certifies that the product identified above have been inspected in compliance with product quality specification and requirements as documented.

Date : 18th April 2023

Authorized Signature

